

Jellies positioning aid is used to alleviate pressure caused by prolonged immobility or other conditions where frequent repositioning is contra-indicated.

## INSTRUCTIONS FOR USE – Single Patient Use Only

1. Read entire Instructions for Use.
2. Warm before using with babies. We recommend placing Jellies in the incubator to allow it to warm. Jellies can be initially warmed by putting it in a plastic bag and placing it under running water, or in your pocket! Once warm, Jellies maintains the temperature of the environment it's in.
3. DO NOT MICROWAVE. Heat spots could result.
4. Cover Jellies with a soft cloth cover or our disposable cover before use. This provides a breathable surface between the baby and the gel pillow.
5. Clean with standard hospital disinfectants.
6. Periodically check for punctures or leakage.
7. Lasts 4 to 6 weeks under normal conditions.
8. For a thicker pillow, push gel to one side, fold under and tape.

## WARNINGS

Jellies are indicated as single patient use to prevent possible cross contamination between patients. Any reuse of Jellies is not warranted by the manufacturer.

1. Hospital use only. Jellies should only be used in a monitored environment.
2. **Do not allow infant to lay face down on pillow – danger of suffocation.**
3. Do not puncture the pillow. If the pillow is punctured replace it. Do not try to repair the puncture.
4. Do not x-ray through pillow. Air bubbles may cause misdiagnosis.
5. Do not store unused Jellies on blanket warmers – can cause uneven hot spots.
6. DO NOT MICROWAVE.

## PRODUCT SPECIFICATIONS

Latex-free. Outer Material: polyolefin film. Contents: Water based non-toxic food grade gel. Methylparaben anti-bacterial agent. See MSDS for more information.

3-1504-002b

Jellies positioning aid is used to alleviate pressure caused by prolonged immobility or other conditions where frequent repositioning is contra-indicated.

## INSTRUCTIONS FOR USE – Single Patient Use Only

1. Read entire Instructions for Use.
2. Warm before using with babies. We recommend placing Jellies in the incubator to allow it to warm. Jellies can be initially warmed by putting it in a plastic bag and placing it under running water, or in your pocket! Once warm, Jellies maintains the temperature of the environment it's in.
3. DO NOT MICROWAVE. Heat spots could result.
4. Cover Jellies with a soft cloth cover or our disposable cover before use. This provides a breathable surface between the baby and the gel pillow.
5. Clean with standard hospital disinfectants.
6. Periodically check for punctures or leakage.
7. Lasts 4 to 6 weeks under normal conditions.
8. For a thicker pillow, push gel to one side, fold under and tape.

## WARNINGS

Jellies are indicated as single patient use to prevent possible cross contamination between patients. Any reuse of Jellies is not warranted by the manufacturer.

1. Hospital use only. Jellies should only be used in a monitored environment.
2. **Do not allow infant to lay face down on pillow – danger of suffocation.**
3. Do not puncture the pillow. If the pillow is punctured replace it. Do not try to repair the puncture.
4. Do not x-ray through pillow. Air bubbles may cause misdiagnosis.
5. Do not store unused Jellies on blanket warmers – can cause uneven hot spots.
6. DO NOT MICROWAVE.

## PRODUCT SPECIFICATIONS

Latex-free. Outer Material: polyolefin film. Contents: Water based non-toxic food grade gel. Methylparaben anti-bacterial agent. See MSDS for more information.

3-1504-002b